



Biosafety Certificate No. 42/13 A

ESCO Airstream Class II (AC2-4E8) Biohazard Safety Cabinet

Tests to BS EN 12469:2000, Biotechnology- Performance
criteria for microbiological safety cabinets

Report No. 42/13 A

Report Prepared For: ESCO
Issue Date: 14th June 2013

Test Summary

The Biosafety Unit of Public Health England has undertaken extensive tests on the ESCO Airstream Class II (AC2-4E8) biohazard safety cabinet. The cabinet was found to meet the performance required by BS EN 12469:2000 for all the tests below

- (A) Air Velocity measurements
- (B) Operator protection/containment tests using KI
- (C) Operator protection tests using bacterial spores
- (D) Product protection tests using bacterial spores
- (E) Cross contamination tests using bacterial spores
- (F) Air visualisation tests
- (G) Filter integrity tests
- (H) Maintenance of temperature by the working cabinet
- (I) Acceptable noise and light levels during operation of the cabinet
- (J) Acceptable design of filter mountings and plenum

Testing Performed By

Name: Mr Simon Parks
Title: Senior Biosafety Scientist

Authorised By

Name: Ms Anna Moy
Title: Biosafety Scientist